

Result of Nuclide Analysis of Seawater <Offshore of Ibaragi Prefecture>

Reference

(Data summarized : July 20)

Place of Sampling	3 km offshore of Takadokobama shore Upper Layer		3 km offshore of Takadokobama shore Lower Layer		3 km offshore of Kujihama shore Upper Layer		3 km offshore of Kujihama shore Lower Layer		3 km offshore of Oarai shore Upper Layer		3 km offshore of Oarai shore Lower Layer		Density limit by the announcement of Reactor Regulation (Bq/ L) (the density limit in the water outside of surrounding monitored areas in the section 6 of the appendix 2)
Time and Date of Sample Collection	July 15, 2011 7:49 am		July 15, 2011 7:47 am		July 16, 2011 8:35 am		July 16, 2011 8:33 am		July 16, 2011 8:12 am		July 16, 2011 8:02 am		
Detected Nuclides (Half-life)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	
I-131 (about 8 days)	ND	-	ND	-	ND	-	ND	-	ND	-	ND	-	40
Cs-134 (about 2 years)	ND	-	ND	-	ND	-	ND	-	ND	-	ND	-	60
Cs-137 (about 30 years)	ND	-	ND	-	ND	-	ND	-	ND	-	ND	-	90

Place of Sampling	3 km offshore of Hirai shore Upper Layer		3 km offshore of Hirai shore Lower Layer		3 km offshore of Hasaki shore Upper Layer		3 km offshore of Hasaki shore Lower Layer						Density limit by the announcement of Reactor Regulation (Bq/ L) (the density limit in the water outside of surrounding monitored areas in the section 6 of the appendix 2)
Time and Date of Sample Collection	July 15, 2011 1:32 pm		July 15, 2011 1:30 pm		July 14, 2011 2:26 pm		July 14, 2011 2:21 pm						
Detected Nuclides (Half-life)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	Density of Sample (Bq/ L)	Scaling Factor (/)	
I-131 (about 8 days)	ND	-	ND	-	ND	-	ND	-	/	/	/	/	40
Cs-134 (about 2 years)	ND	-	ND	-	ND	-	ND	-	/	/	/	/	60
Cs-137 (about 30 years)	ND	-	ND	-	ND	-	ND	-	/	/	/	/	90

Density by the announcement of Reactor Regulation is stated with an amount converted from Bq/cm³ to Bq/L

Data of other nuclides are under evaluation.

In the case that two or more kinds of nuclides exist, sum of each scaling factor to the density limit is compared with 1.

In the case that the data is below measurable limit, "ND" is stated.

Detection limits of the three main nuclides are as follows: I-131: approx. 6Bq/L, Cs-134: approx. 9Bq/L, Cs-137: approx. 9Bq/L

Please note that these nuclides are sometimes detected even when they are below the threshold, contingent on the detector or samples.